



APPENDIX A
ACCUTORR V GUIDELINE

(This document is a guideline only to be used as an aid to comprehensive Inservice training.)

NAME: _____

HOSPITAL: _____

DATE: _____

VALIDATOR: _____

	Performed	Not Performed	Not Applicable
A. General			
1. Locate on/standby key			
2. Locate main power switch and battery operation indicator			
3. Locate the trend window			
4. Locate the function keys			
5. Admit a patient (Bar code scanner) Locate the current patient information			
6. Locate and load recorder paper			
B. Basic Functions			
1. Initiate an SpO ₂ reading			
2. Discuss SpO ₂ sensor application for reusable and disposable sensors			
3. Print the SpO ₂ waveform			
4. Locate HR display and Identify HR sources			
5. Select the Adult patient size			
6. Initiate a Non-Invasive Blood Pressure (NIBP) reading			
7. Stop an NIBP reading			

	Performed	Not Performed	Not Applicable
8. Set the NIBP interval for every 30 minutes			
9. Pause the NIBP interval How does the clinician know the NIBP interval is interrupted?			
10. Resume the NIBP interval measurements			
11. Set the High SYS alarm limit to 200			
12. Set the Low HR alarm to 50			
13. Silence an alarm a. for 2 minutes b. Until a new alarm occurs			
14. Print: a. Last 5 trends on the patient b. All trends on the patient c. Stop the printing			
15. Delete the most recent trended information			
16. Delete trended data for: a. this patient b. all patients			
17. Reset the monitor to "preset" (default) settings			
18. Obtain oral temperature			
19. Adjust temp unit to °C			
C. Advance Functions			
1. Reset the monitor to "user config settings" (default)			
2. Change the date/time			
3. Change the default settings to "turn on" in Pediatrics			
4. Adjust the user configuration settings			