



**DPM Central Station  
Default Settings**

Hospital: \_\_\_\_\_

Department: \_\_\_\_\_

Clinical / Sales / Service Representative: \_\_\_\_\_

(\* Optional)

Date: \_\_\_\_\_ Software Version: \_\_\_\_\_

(\* \*\* Teleanck)

Signature of Departmental Representative: \_\_\_\_\_ Title: \_\_\_\_\_

Recorder Control			Options	Selection
	Recording Duration		8, 16, 32 sec	
	Sweep Speed		25, 50 mm/s	
	Grid		On, Off	
Printina Control			Options	Selection
	Printer		HP 1505 N, HP 2035 N, HP 2055DN, HP 4015 N	
General Setup			Options	Selection
	Volume Levels		1-10	
	Alarm Off Reminder Volume		Low, Med, High	
	Telemetry Alarm Delay Setup**	Alarm Delay	Off, 1, 2, 3, 4, 5, 6, 7, 8 sec	
		ST alarm delay	30, 45 sec; 1, 1.5, 2, 3 min	
	Waveform Rendering		Color, Mono	
		CO2*	mmHg, kPa	
		Temp	C, F	
		Vascular Resistance	DS/cm <sup>5</sup> , kPa-s/l	
		IBP*	mmHg, kPa, cmH2O	
		Gas CO2*	mmHg, kPa, %	
		Gas O2*	mmHg, kPa, %	
		ECG ST	mV, mm	
Admin Setup (Password Required)			Options	Selection
Color	Parameter	ECG	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		Resp	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		SPO2	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		Temp	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	



		LAP*	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		BIS*	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		RM*	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
<b>Admin Setup (Password Required)</b>			<b>Options</b>	<b>Selection</b>
<b>Color</b>	<b>Parameter</b>	ICG*	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		LAP*	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
	<b>Arrhythmia (Arr.)</b>	Asystole	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		VFib	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		Vlac	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		PVCs/min	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		R on T	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		VT>2	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		Couplet	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		Multif. PVC	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		Vent Rhythm	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		Bigeminy	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		Trigeminy	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		Tachy	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		Brady	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		Irr. Rhythm	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		PNP	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		PNC	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
		Missed Beats	White, Gray 1-3, Red, Red 1-3, Orange, Orange 1-3, Yellow, Yellow 1-3, Green, Green 1-3, Cyan, Cyan 1-3, Blue, Blue 1-3, Purple, Purple 1-3	
<b>Display Setup</b>	<b>Screen Size</b>		15", 17", 19", 20", 21"	
	<b>Multibed Display</b>	Measurement Units	Show, Hide	
		Patient Name	Show, Hide	
	<b>Name Displaying Hotkey</b>		Off, F1, F2, F3, F4, F5, F6, F7, F8, F9, F10	
	<b>Layout (max 16 per display)</b>	Rows	2, 3, 4, 5, 6, 7, 8	
		Columns	1, 2, 3	
	<b>Auxiliary Window Setup</b>	1st Tab	Patient management, ViewBed, Waveform Review, Trend Review, Alarm Review, Display Setup, ST Review, Drug Calculation, Hemodynamic Calculation, Oxygenation Calculation, Renal Calculation, Ventilation Calculation	
		2nd Tab	Patient management, ViewBed, Waveform Review, Trend Review, Alarm Review, Display Setup, ST Review, Drug Calculation, Hemodynamic Calculation, Oxygenation Calculation, Renal Calculation, Ventilation Calculation	

		3rd Tab	Patient management, ViewBed, Waveform Review, Trend Review, Alarm Review, Display Setup, ST Review, Drug Calculation, Hemodynamic Calculation, Oxygenation Calculation, Renal Calculation, Ventilation Calculation	
		4th Tab	Patient management, ViewBed, Waveform Review, Trend Review, Alarm Review, Display Setup, ST Review, Drug Calculation, Hemodynamic Calculation, Oxygenation Calculation, Renal Calculation, Ventilation Calculation	
		5th Tab	Patient management, ViewBed, Waveform Review, Trend Review, Alarm Review, Display Setup, ST Review, Drug Calculation, Hemodynamic Calculation, Oxygenation Calculation, Renal Calculation, Ventilation Calculation	
		6th Tab	Patient management, ViewBed, Waveform Review, Trend Review, Alarm Review, Display Setup, ST Review, Drug Calculation, Hemodynamic Calculation, Oxygenation Calculation, Renal Calculation, Ventilation Calculation	
		7th Tab	Patient management, ViewBed, Waveform Review, Trend Review, Alarm Review, Display Setup, ST Review, Drug Calculation, Hemodynamic Calculation, Oxygenation Calculation, Renal Calculation, Ventilation Calculation	
<b>Admin Setup (Password Required)</b>			<b>Options</b>	<b>Selection</b>
	<b>Auxiliary Window Setup</b>	8th Tab	Patient management, ViewBed, Waveform Review, Trend Review, Alarm Review, Display Setup, ST Review, Drug Calculation, Hemodynamic Calculation, Oxygenation Calculation, Renal Calculation, Ventilation Calculation	
		9th Tab	Patient management, ViewBed, Waveform Review, Trend Review, Alarm Review, Display Setup, ST Review, Drug Calculation, Hemodynamic Calculation, Oxygenation Calculation, Renal Calculation, Ventilation Calculation	
		10th Tab	Patient management, ViewBed, Waveform Review, Trend Review, Alarm Review, Display Setup, ST Review, Drug Calculation, Hemodynamic Calculation, Oxygenation Calculation, Renal Calculation, Ventilation Calculation	
		11th Tab	Patient management, ViewBed, Waveform Review, Trend Review, Alarm Review, Display Setup, ST Review, Drug Calculation, Hemodynamic Calculation, Oxygenation Calculation, Renal Calculation, Ventilation Calculation	
		12th Tab	Patient management, ViewBed, Waveform Review, Trend Review, Alarm Review, Display Setup, ST Review, Drug Calculation, Hemodynamic Calculation, Oxygenation Calculation, Renal Calculation, Ventilation Calculation	
<b>Alarm</b>		Alarm Audio Off	On, Off	
		Remote Alarm Control	On, Off	
		Alarm Latched	On, Off	
		Alarm Off Reminder Tone	On, Off	
		Reminder Interval	1, 2, 3 min	
		Silence Hotkey	Off, F1, F2, F3, F4, F5, F6, F7, F8, F9, F10	
		Alarm Audio Mode	ISO, Mode 1, Mode 2	
	<b>Alarm Sound Intervals (ISO)</b>	High Level	3-15 sec	
		Medium Level	3-20 sec	
		Low Level	15-100 sec	
		Alarm Off Reminder Interval	1, 2, 3 min	
	<b>Alarm Numeric and Related Waveforms</b>	HR	ECG 1- ECG 7*, Pleth, IBP1-4*, ART*, PA*, CVP*, RAP*, LAP*, ICP*, Ao*, UAP*, BAP*, FAP*, UVP*, LV*, P1-4*, Resp, CO2*, Gas CO2*, Gas O2*, Gas N2O*, Gas AA*, BIS*, Flow*, Paw*, Vol*, ICG*	
		ECG Lost	ECG 1- ECG 7*, Pleth, IBP1-4*, ART*, PA*, CVP*, RAP*, LAP*, ICP*, Ao*, UAP*, BAP*, FAP*, UVP*, LV*, P1-4*, Resp, CO2*, Gas CO2*, Gas O2*, Gas N2O*, Gas AA*, BIS*, Flow*, Paw*, Vol*, ICG*	
		PVC	ECG 1- ECG 7*, Pleth, IBP1-4*, ART*, PA*, CVP*, RAP*, LAP*, ICP*, Ao*, UAP*, BAP*, FAP*, UVP*, LV*, P1-4*, Resp, CO2*, Gas CO2*, Gas O2*, Gas N2O*, Gas AA*, BIS*, Flow*, Paw*, Vol*, ICG*	
		ST1*	ECG 1- ECG 7*, Pleth, IBP1-4*, ART*, PA*, CVP*, RAP*, LAP*, ICP*, Ao*, UAP*, BAP*, FAP*, UVP*, LV*, P1-4*, Resp, CO2*, Gas CO2*, Gas O2*, Gas N2O*, Gas AA*, BIS*, Flow*, Paw*, Vol*, ICG*	
		ST2*	ECG 1- ECG 7*, Pleth, IBP1-4*, ART*, PA*, CVP*, RAP*, LAP*, ICP*, Ao*, UAP*, BAP*, FAP*, UVP*, LV*, P1-4*, Resp, CO2*, Gas CO2*, Gas O2*, Gas N2O*, Gas AA*, BIS*, Flow*, Paw*, Vol*, ICG*	
		RR/AgAwRR*/RM RR*	ECG 1- ECG 7*, Pleth, IBP1-4*, ART*, PA*, CVP*, RAP*, LAP*, ICP*, Ao*, UAP*, BAP*, FAP*, UVP*, LV*, P1-4*, Resp, CO2*, Gas CO2*, Gas O2*, Gas N2O*, Gas AA*, BIS*, Flow*, Paw*, Vol*, ICG*	
		Resp Apnea/CO2 Apnea*/Gas Apnea*, RM Apnea*	ECG 1- ECG 7*, Pleth, IBP1-4*, ART*, PA*, CVP*, RAP*, LAP*, ICP*, Ao*, UAP*, BAP*, FAP*, UVP*, LV*, P1-4*, Resp, CO2*, Gas CO2*, Gas O2*, Gas N2O*, Gas AA*, BIS*, Flow*, Paw*, Vol*, ICG*	
		Resp Artifact	ECG 1- ECG 7*, Pleth, IBP1-4*, ART*, PA*, CVP*, RAP*, LAP*, ICP*, Ao*, UAP*, BAP*, FAP*, UVP*, LV*, P1-4*, Resp, CO2*, Gas CO2*, Gas O2*, Gas N2O*, Gas AA*, BIS*, Flow*, Paw*, Vol*, ICG*	
		SPO2	ECG 1- ECG 7*, Pleth, IBP1-4*, ART*, PA*, CVP*, RAP*, LAP*, ICP*, Ao*, UAP*, BAP*, FAP*, UVP*, LV*, P1-4*, Resp, CO2*, Gas CO2*, Gas O2*, Gas N2O*, Gas AA*, BIS*, Flow*, Paw*, Vol*, ICG*	





			HR, SPO2, NIBP, RR, PR, IBP1*, IBP2*, IBP3*, IBP4*, ART*, CVP*, PA*, Ao*, UAP*, BAP*, FAP*, ICP*, LAP*, RAP*, UVP*, LV*, P1*, P2*, P3*, P4*, T1, T2, ST1*, ST2*, ST I*, ST II*, AWRR*, I/E CO2*, I/E O2*, I/E N2O*, I/E Agents *(AA, HAL, ENF, ISO, SEV, DES), AG AWRR*, TB, RMRR*, PEEP*, PIP*, MVe*, Pplat*, Tve, BIS*	
	<b>Trend Group (Order)</b>	Custom:		
		Custom:	HR, SPO2, NIBP, RR, PR, IBP1*, IBP2*, IBP3*, IBP4*, ART*, CVP*, PA*, Ao*, UAP*, BAP*, FAP*, ICP*, LAP*, RAP*, UVP*, LV*, P1*, P2*, P3*, P4*, T1, T2, ST1*, ST2*, ST I*, ST II*, AWRR*, I/E CO2*, I/E O2*, I/E N2O*, I/E Agents *(AA, HAL, ENF, ISO, SEV, DES), AG AWRR*, TB, RMRR*, PEEP*, PIP*, MVe*, Pplat*, Tve, BIS*	
		Custom:	HR, SPO2, NIBP, RR, PR, IBP1*, IBP2*, IBP3*, IBP4*, ART*, CVP*, PA*, Ao*, UAP*, BAP*, FAP*, ICP*, LAP*, RAP*, UVP*, LV*, P1*, P2*, P3*, P4*, T1, T2, ST1*, ST2*, ST I*, ST II*, AWRR*, I/E CO2*, I/E O2*, I/E N2O*, I/E Agents *(AA, HAL, ENF, ISO, SEV, DES), AG AWRR*, TB, RMRR*, PEEP*, PIP*, MVe*, Pplat*, Tve, BIS*	
		Custom:	HR, SPO2, NIBP, RR, PR, IBP1*, IBP2*, IBP3*, IBP4*, ART*, CVP*, PA*, Ao*, UAP*, BAP*, FAP*, ICP*, LAP*, RAP*, UVP*, LV*, P1*, P2*, P3*, P4*, T1, T2, ST1*, ST2*, ST I*, ST II*, AWRR*, I/E CO2*, I/E O2*, I/E N2O*, I/E Agents *(AA, HAL, ENF, ISO, SEV, DES), AG AWRR*, TB, RMRR*, PEEP*, PIP*, MVe*, Pplat*, Tve, BIS*	
		Custom:	HR, SPO2, NIBP, RR, PR, IBP1*, IBP2*, IBP3*, IBP4*, ART*, CVP*, PA*, Ao*, UAP*, BAP*, FAP*, ICP*, LAP*, RAP*, UVP*, LV*, P1*, P2*, P3*, P4*, T1, T2, ST1*, ST2*, ST I*, ST II*, AWRR*, I/E CO2*, I/E O2*, I/E N2O*, I/E Agents *(AA, HAL, ENF, ISO, SEV, DES), AG AWRR*, TB, RMRR*, PEEP*, PIP*, MVe*, Pplat*, Tve, BIS*	
		Custom:	HR, SPO2, NIBP, RR, PR, IBP1*, IBP2*, IBP3*, IBP4*, ART*, CVP*, PA*, Ao*, UAP*, BAP*, FAP*, ICP*, LAP*, RAP*, UVP*, LV*, P1*, P2*, P3*, P4*, T1, T2, ST1*, ST2*, ST I*, ST II*, AWRR*, I/E CO2*, I/E O2*, I/E N2O*, I/E Agents *(AA, HAL, ENF, ISO, SEV, DES), AG AWRR*, TB, RMRR*, PEEP*, PIP*, MVe*, Pplat*, Tve, BIS*	
<b>Admin Setup (Password Required)</b>			<b>Options</b>	<b>Selection</b>
<b>Telemetry</b>	<b>System Information</b>	Main Board Software Version		
		Receiver board Software Version		
	<b>Alarm Level Setup</b>	Alm Lev of ECG Lead Off	Low, Med, High	
		Alm Lev of SPO2 Sensor Off	Low, Med, High	
	<b>Monitoring Mode Setup</b>		Fixed Monitoring, Flexible Monitoring	
<b>Other</b>	<b>Change User Password</b>			
	<b>Change Hospital Name</b>	Hospital Name		
		Unit Name		
<b>Factory Setup (Password Required)</b>			<b>Options</b>	<b>Selection</b>
	<b>Hours of Stored Waveforms</b>		72, 240 hours	
	<b>Version</b>			
<b>Windows Setup (Control Panel Date/Time Configuration)</b>			<b>Options</b>	<b>Selection</b>
	<b>Hour Format</b>		hmmssst, hhmmssst, HHmmss, Hmss	
	<b>Date Format</b>		Short date format, Long date format (including format configuration)	

Parameter Setup				Options						Selection		
HR	HR Activation State			Refer to Alarm Setup Menu								
	HR Alarm Priority			Refer to Alarm Setup Menu								
	HR Limit			Refer to Alarm Setup Menu								
	Pacer Rate			40-100								
	Sweep Speed			6.25, 12.5, 25, 50 mm/s								
	HR Source			ECG, SPO2, Auto, Both								
	Waveform Setup	Filter		Monitor, Surgical, Extended								
		Channel 1 Waveform Gain		0.125, 0.25, 0.5, 1, 2, 4								
		Channel 2 Waveform Gain		0.125, 0.25, 0.5, 1, 2, 4								
		Channel 3 Waveform Gain		0.125, 0.25, 0.5, 1, 2, 4								
ST*	ST Analysis			On, Off								
	ST Activation State			Refer to Alarm Setup Menu								
	ST Alarm Priority			Refer to Alarm Setup Menu								
	ST Alarm Setup			mmHg, kPa								
Arrhythmia	Arrhythmia Analysis			On, Off								
Parameter Alarm Settings		High		Low		Activation State		Alarm Priority			Record On Alarm	
	Heart Rate	17 - 300 bpm		15 - 298 bpm		ON	OFF	High	Med	Low	ON	OFF
	ST-I*	-18 - +20 mm; -1.8 - +2mV		-18 - +20 mm; -1.8 - +2mV		ON	OFF	High	Med	Low	ON	OFF
	ST-II*	-18 - +20 mm; -1.8 - +2mV		-18 - +20 mm; -1.8 - +2mV		ON	OFF	High	Med	Low	ON	OFF
	ST-III*	-18 - +20 mm; -1.8 - +2mV		-18 - +20 mm; -1.8 - +2mV		ON	OFF	High	Med	Low	ON	OFF
	ST-aVR*	-18 - +20 mm; -1.8 - +2mV		-18 - +20 mm; -1.8 - +2mV		ON	OFF	High	Med	Low	ON	OFF
	ST-aVL*	-18 - +20 mm; -1.8 - +2mV		-18 - +20 mm; -1.8 - +2mV		ON	OFF	High	Med	Low	ON	OFF
	ST-aVF*	-18 - +20 mm; -1.8 - +2mV		-18 - +20 mm; -1.8 - +2mV		ON	OFF	High	Med	Low	ON	OFF
	ST-V*	-18 - +20 mm; -1.8 - +2mV		-18 - +20 mm; -1.8 - +2mV		ON	OFF	High	Med	Low	ON	OFF
	SpO2	1 - 100%		0 - 99%		ON	OFF	High	Med	Low	ON	OFF
	PR	20 - 300 bpm		18 - 298 bpm		ON	OFF	High	Med	Low	ON	OFF



Arrhythmia Alarm		Activation State					Record On Alarm	
	Asystole	ON	OFF	High			ON	OFF
	Vfib	ON	OFF	High			ON	OFF
	Vtac	ON	OFF	High			ON	OFF
	PVCs/min	ON	OFF	High	Med	Low	ON	OFF
	R on T	ON	OFF	High	Med	Low	ON	OFF
	VT>2	ON	OFF	High	Med	Low	ON	OFF
	Couplet	ON	OFF	High	Med	Low	ON	OFF
	Multif PVC	ON	OFF	High	Med	Low	ON	OFF
	Vent. Rhythm	ON	OFF	High	Med	Low	ON	OFF
	Bigeminy	ON	OFF	High	Med	Low	ON	OFF
	Trigeminy	ON	OFF	High	Med	Low	ON	OFF
	Tachy	ON	OFF	High	Med	Low	ON	OFF
	Brady	ON	OFF	High	Med	Low	ON	OFF
	Irr. Rhythm	ON	OFF	High	Med	Low	ON	OFF
	PNP	ON	OFF	High	Med	Low	ON	OFF
	PNC	ON	OFF	High	Med	Low	ON	OFF
	Missed Beat	ON	OFF	High	Med	Low	ON	OFF
Arrhythmia Threshold		Options					Selection	
	PVCs High	3 - 31 PVC						
	Asys Delay	3 - 10 sec						
	V Tac Rate	100 - 200 bpm						
	V Tac PVC	3 - 12 PVC						
	Multi PVCs Window	3 - 31 PVC						
	Tachy High	100 - 300 bpm						
	Brady Low	15 - 60 bpm						